

For intraoral radiography a simple test tool – an image of a step wedge is useful for maintaining high image quality (Fig.8) . During installation, a reference standard radiograph of the step wedge should be made using the optimized exposure setting for an adult/child. Subsequent radiographs of the step wedge/phantom should be made during clinical use and compared with the reference one to ensure that image quality is maintained [EC-RP 136]

